Molecular typing for the Indian blood group associated 252G>C single nucleotide polymorphism in a selected cohort of Australian blood donors

Genghis H. Lopez¹, Rhiannon S. McBean¹, Brett Wilson², Darryl L. Irwin³, Yew-Wah Liew², Catherine A. Hyland¹, Robert L. Flower¹

¹Research and Development Division; ²Red Cell Reference Laboratory, Australian Red Cross Blood Service, Queensland, Australia; ³Sequenom Inc. Asia Pacific, Herston, Queensland, Australia

Background. The Indian blood group antigens, In^a and In^b , are clinically significant in transfusion medicine. However, antisera to type these antigens are difficult to obtain. The In^b antigen is a high frequency antigen present in all populations, while the frequency of the antithetical In^a ranges from 0.1% in Caucasians up to 11% in Middle Eastern groups. This antigen polymorphism is encoded by the single nucleotide polymorphism (SNP) 252G>C in *CD44*. The aim of this study was to establish and compare two genotyping methods to measure the frequency of the *IN*A* and *IN*B* alleles in a blood donor cohort.

Materials and methods. Donor blood samples (n=151) were genotyped by a novel real-time polymerase chain reaction (PCR) high-resolution meltcurve (HRM) analysis and a custom matrix-assisted laser desorption/ionisation time-of-flight mass spectrometry (MALDI-TOF MS) assay. Samples with the rare IN*A allele were further investigated by nucleotide sequencing, red cell agglutination, and flow cytometry techniques.

Results. In this study group, 149 IN*B homozygous and 2 IN*A/B heterozygous samples were detected with 100% concordance between HRM and MALDI-TOF MS methods. For PCR HRM, amplicon melting alone did not differentiate IN*A and IN*B alleles (class 3 SNP), however, the introduction of an unlabelled probe (UP) increased the resolution of the assay. Sequencing confirmed that the two non-homozygous samples were IN*A/B heterozygous and phenotyping by red cell agglutination, and flow cytometry confirmed both In^a and In^b antigens were present as predicted.

Discussion. Genotyping permits conservation of rare antisera to predict blood group antigen phenotype. In PCR UP-HRM the IN*A and IN*B alleles were discriminated on the basis of their melting properties. The In^a frequency in this selected donor population was 1.3%. Application of genotyping methods such as these assists in identifying donors with rare blood group phenotypes of potential clinical significance.

Keywords: Indian blood group genotyping, high-resolution melting analysis, MALDI-TOF MS.

Introduction

The *CD44* gene, located on chromosome 11p13 consists of 20 exons, encoding numerous isoforms of the CD44 molecule^{1,2}. The CD44 isoform present on hematopoietic cells is CD44s^{1,3} and approximately 2,000-5,000 copies of this molecule are expressed on mature red blood cells⁴. The CD44 glycoprotein carries antigens of the Indian (IN) blood group system⁵, namely: In^a, In^b, INFI, and INJA⁶⁻⁹. The In^a and In^b are antithetical antigens and have been associated with a single nucleotide polymorphism (SNP) from guanine (encoding In^b) to cytosine (encoding In^a) at nucleotide 252 in exon 2 of *CD44*. This SNP results in an amino acid substitution Arg46Pro of the CD44 molecule¹⁰.

The In^a and In^b antigens are efficient immunogens and antibodies to these polymorphisms are clinically significant¹¹⁻¹⁴. Anti-In^a and anti-In^b antibodies have been associated with immediate and delayed haemolytic transfusion reactions^{7,13,15}. The In^b antigen is a high frequency antigen expressed in all populations. While In^a is expressed in 0.1% of Caucasians⁴, it is present in 2.9% of Indians, 10.6% of Iranians, and 11.8% of Arab populations living in Mumbai (Bombay, India)^{7,16}. These studies on the population frequency of In^a and In^b were performed over 30 years ago^{7,11,16,17}.

The antisera required for typing the In^a and In^b antigens are rare, and difficult to obtain. The SNP associated with the In^a and In^b antigens has been defined; however, commercially-available genotyping systems such as BLOODchip v4.1 (Progenika, Vizcaya, Spain) and Human Erythrocyte Antigen (HEA) BeadChip Kit (BioArray Solutions Ltd, Warren, NJ, USA) do not incorporate this polymorphism. The aim of this study was to apply two DNA methods to genotype the *IN*A* and *IN*B* alleles and to investigate the frequency of In^a and In^b in a selected population of Australian blood donors. We describe a novel high-resolution meltcurve (HRM) analysis method and a custom matrix-assisted laser desorption/ionisation time-of-flight mass spectrometry (MALDI-TOF MS) assay to genotype the *IN*A* and *IN*B* alleles. Molecular typing results were validated by comparing data obtained between genotyping methods, with phenotype determined by standard serological techniques, and Sanger DNA sequencing.

Materials and methods Donor selection and DNA preparation

Blood samples collected into EDTA (n=151) were retrieved from volunteer blood donors who presented at collection sites where the population includes South Asian and West Asian (Middle Eastern) individuals. This study had ethics approval from the Human Research Ethics Committee, Australian Red Cross Blood Service.

Two additional samples, In(a+b-) and In(a-b+), were used as controls in genotyping experiments. The In(a+b-) and In(a-b+) control samples were phenotyped by the Blood Service Red Cell Reference Laboratories. The In(a+b-) control was kindly provided by the Sydney Red Cell Reference Laboratory.

Genomic DNA was extracted from whole blood using the EZ1 DNA Blood kit in an automated BioRobot EZ1 Workstation (QIAGEN, Doncaster, VIC, Australia) in accordance with the manufacturer's instructions.

Polymerase chain reaction unlabelled probe highresolution meltcurve analysis

HRM analysis is a post-polymerase chain reaction (PCR), DNA analysis method used for genotyping and scanning DNA sequence variants¹⁸. HRM uses intercalating dyes that fluoresce when bound to double-stranded (ds) DNA only. Following PCR, dsDNA is melted by gradually increasing the temperature. As dsDNA dissociates into single-stranded products, total fluorescence is reduced and monitored over time¹⁹. The standard HRM method is performed by amplicon melting only. A variation of this method is the addition of an

unlabelled probe to improve specificity of the assay to detect the target polymorphism during denaturation²⁰⁻²³.

The primers and probe used in this method were designed in reference to GenBank accession number M59040. Oligonucleotide sequences are listed in Table I. The Indian HRM assay primers amplify a 104-bp product. The UP is a complete match to IN*B and is therefore a single-base mismatch to IN*A. It has an amino-C6 modification on the 3'-end to prevent polymerase extension during PCR²⁴. The primers and probe were ordered from Sigma-Aldrich (Sigma-Aldrich, Castle Hill, NSW, Australia).

The PCR reaction volume was 20 μ L and contained 150 nM forward primer (limiting), 750 nM reverse primer (excess), 750 nM unlabelled probe, 20 ng of genomic DNA, and 12.5 μ L of HRM PCR master mix (QIAGEN Type-it HRM PCR kit). This assay utilises an asymmetric PCR technique resulting in the generation of excess single-stranded template available for hybridisation of the probe.

The PCR was performed with an initial activation step of 95 °C for 5 minutes followed by 40 cycles of 95 °C for 10 seconds and an annealing/extension step of 60 °C for 30 seconds. Fluorescence data were acquired on the Green Channel during PCR. After the last cycle, samples were held at 95 °C for 1 minute prior to lowering the temperature to 65 °C. Melting was then performed by steadily raising the temperature at increments of 0.5 °C every 2 seconds from 65 °C to 90 °C. The rate of change in fluorescence was monitored on the HRM channel.

Both PCR and HRM steps were performed using real-time PCR equipment, Rotor-Gene 6000 and Rotor-Gene Q, 5plex HRM platform (QIAGEN). Rotor-Gene Q Series Software was used for data analysis. A confidence threshold was set at 90% for genotype call assignment.

Matrix-assisted laser desorption/ionisation time-offlight mass spectrometry

MassARRAY (Sequenom Inc., San Diego, CA, USA) is a SNP genotyping platform which utilises MALDI-TOF MS technology. SNP genotyping using MassARRAY was performed following the manufacturer's recommendations and involves target-specific PCR amplification followed by target-specific primer single base extension²⁵.

Table I - Primer and probe sequences used in molecular typing for 252C>G associated with IN*A/B.

Method	Oligonucleotide	Sequence	
	IN-Ex2 forward primer	5'-CGCTTTGCAGGTGTATTCC-3'	
HRM assay	IN-Ex2 reverse primer	5'-ATTGTGGGCAAGGTGCTATT-3'	
	IN-33F probe	5'-ATGGTCGCTACAGCATCTCTC <u>G</u> GACGGAGGCCG-3'	
	Forward primer9	5'- TGTTAACCAGGCTGGTCTTGAG-3'	
Exon 2 sequencing	Reverse primer9	5'- AGTTCTAAGCCCAGCTGCCTG-3'	

The nucleotide base underlined in the IN33-F probe sequence is a complete match to IN*B and a single mismatch to IN*A.

MassARRAY provides customisation via the design of amplification and extension primers specific for the SNP of interest^{26,27} using Assay Design 4 software. During the primer extension stage an oligonucleotide primer anneals immediately adjacent to the SNP being genotyped²⁵ and is extended a single base, incorporating the nucleotide at the SNP position. During MALDI-TOF MS, samples are individually irradiated with a short laser pulse which causes the matrix and DNA fragments to desorb, ionise and be accelerated in an electric field²⁵. The time-of-flight and mass of each extension primer is calculated from the time of arrival at the detector, allowing the nucleotide at the SNP to be determined²⁵ in an automated fashion using Typer 4.0 software.

In this study, SNP genotyping was performed following the Sequenom MassARRAY iPLEX Pro genotyping protocol. Primer sequences are available upon request from Sequenom, Inc.

Preparation of the polymerase chain reaction product for DNA sequencing

Primers published by Poole *et al.*⁹ targeting exon 2 of *CD44* were used to amplify a 430-bp product. The FastStart High Fidelity PCR System, dNTPack (Roche Diagnostics, Castle Hill, NSW, Australia) PCR kit was used to make up 50 μ L of master mix according to the manufacturer's recommendation.

Briefly, the PCR started at 95 °C for 2 minutes followed by 35 cycles at 95 °C for 30 seconds, 65 °C for 30 seconds, 72 °C for 1 minute, and a final extension step at 72 °C for 7 minutes on a Veriti Dx Thermal Cycler (Applied Biosystems, Life Technologies, VIC, Australia). PCR products were purified using a MinElute PCR Purification Kit (QIAGEN) according to the manufacturer's protocol. Purified PCR products were sent to the Australian Genome Research Facility (AGRF), The University of Queensland, QLD, Australia, for DNA sequencing. Data were analysed using Chromas Lite 2.01 software.

Low ionic strength solution-indirect antiglobulin testing

Red cells of all 151 samples were washed in phosphate-buffered saline and resuspended in low ionic strength solution (LISS). Samples were tested in DiaMed Card LISS/Coombs (Bio-Rad, NSW, Australia) using a standard protocol for indirect antiglobulin testing (IAT). Samples were incubated with anti-In^b at 37 °C for 15 minutes and then centrifuged using the DiaMed-ID Micro Typing System (Bio-Rad) at 910 rpm for 10 minutes. Samples were examined for agglutination. The anti-In^b human polyclonal antiserum was from the Serum, Cells, And Rare Fluids (SCARF) exchange programme. A limited supply of anti-In^a human polyclonal antiserum was used to perform standard LISS-IAT, as described above, on samples found to possess the *IN*A* allele. Anti-In^a antiserum was kindly provided by Joyce Poole and Nicole Thorton of the International Blood Group Reference Laboratory, Bristol, UK.

Flow cytometry

A flow cytometry experiment was performed on samples positive for the In^a antigen by LISS-IAT. Cell suspensions were incubated with anti-In^a and anti-In^b (the same antisera were used in the LISS-IAT) in separate tubes at 37 °C for 30 minutes. Cells were washed in phosphatebuffered saline and incubated with PE Mouse Anti-Human IgG (Cat No. 555787, BD Biosciences, Franklin Lakes, NJ, USA) at room temperature in the dark for 30 minutes. Following secondary antibody incubation, cells were washed and resuspended in phosphate-buffered saline. Samples were analysed using a FACS CANTO II flow cytometer and FACSDiva software (BD Biosciences).

Results

Development of the polymerase chain reaction unlabelled probe high-resolution meltcurve assay

The PCR products derived from IN^*A/A and IN^*B/B control samples were denatured between 65.0 °C and 90.0 °C, and generated two distinct domains: an amplicon melting domain from 80.5 °C to 87.0 °C and a probe melting domain from 69.0 °C to 80.5 °C (Figure 1a).

In the amplicon melting domain, the two control samples produced identical melting temperatures (T_m) at 84.0 °C. Melt profiles in this domain showed indistinguishable melt profiles masking the single nucleotide difference between IN^*A and IN^*B . Data in this domain were unsuitable for genotyping analysis.

In the probe melting domain, control samples displayed distinct melting behaviour. The $T_{\rm m}$ was 73.4 °C for the *IN*A/A* control and 77.0 °C for the *IN*B/B* control, a temperature shift of 3.6 °C. These unique melt patterns supported the assignment of homozygous *IN*A/A* and *IN*B/B* genotypes to the two control samples In(a+b-) and In(a-b+) respectively. These melt profiles were then used to assign genotype to our sample cohort.

Polymerase chain reaction unlabelled probe highresolution meltcurve genotyping analysis

The melt profiles of 149 (149/151) unknown samples matched the *IN*B/B* genotype profile with a percentage confidence of at least 98% (Figure 1b). Two samples (2/151) produced peaks at 73.3 °C and 76.6 °C, in the probe T_m melting domain of melting curves for *IN*A* and *IN*B*, respectively (Table II).



Figure 1a - Representative melt profiles of samples homozygous for IN^*A/A (thick line) and IN^*B/B (thin line). Meltcurve graph plots the rate of change in fluorescence (y-axis) against temperature (x-axis).



Figure 1b - Representative melt profile of IN^*A/B heterozygous sample (dash) produced double peaks in the probe melting domain at 73.3±0.3 °C and 76.6±0.2 °C compared to the single peaks produced by IN^*A (T_m =73.4 °C) and IN^*B (T_m =77.0 °C) homozygous samples.

Table II -	Average $T_{\rm m}$ of homozygous IN^*A/A and IN^*B/A		
	and donor cells: mean $T_{\rm m}$ and two standard		
	deviations (SD) were calculated from 30		
	replicate assays.		

Sample	Melting domain T _m Probe±2SD	Melting domain T _m Amplicon±2SD
Control IN*A/A	73.4±0.3 °C	84.0±0.3 °C
Control IN*B/B	77.0±0.3 °C	84.0±0.3 °C
Donor 1 IN*A/B	73.3±0.3 °C and 76.6±0.2 °C	83.8±0.2 °C
Donor 2 IN*A/B	73.3±0.3 °C and 76.6±0.3 °C	83.8±0.2 °C

Blood Transfus 2015; 13: 78-85 DOI 10.2450/2014.0336-13

Matrix-assisted laser desorption/ionisation timeof-flight mass spectrometry genotyping analysis

On analysis the In(a+b-) control produced a single peak with a mass of 5328.5 Dalton (Da) consistent with the cytosine SNP (Figure 2a). The In(a-b+) control produced a single peak of 5368.5 Da consistent with the guanine SNP (Figure 2b). One hundred and forty-nine samples produced single peaks with a mass of 5368.5 Da and were therefore genotyped as homozygous for guanine (IN*B/B). Two samples produced peaks at both 5328.5 Da and 5368.5 Da indicating they were heterozygous for cytosine and guanine (IN*A/B) (Figure 2c). MALDI-TOF MS genotyping results of the control samples and all 151 unknown samples were fully concordant with genotypes determined by unlabelled probe (UP)-HRM.

DNA sequencing

Sequencing data of the two homozygous controls and the two identified IN^*A/B heterozygous samples were aligned with the reference allele M59040 using the Basic Local Alignment Tool (BLAST) on the NCBI database website (http://www.ncbi.nlm.nih. gov). Chromatograms (Figure 3) show that a cytosine base for IN*A/A was substituted by a guanine base in IN*B/B, whereas for the two heterozygous IN*A/B samples, both cytosine and guanine were detected in the same position.



a - MALDI-TOF MS spectrum of control IN*A/A. A single peak at 5328.5 Da is consistent with the cytosine SNP.



b - MALDI-TOF MS spectrum of control IN*B/B. A single peak at 5368.5 Da is consistent with the guarine SNP.



c - MALDI-TOF MS spectrum of control *IN*A/B* samples. Heterozygous samples produce peaks at 5328.5 Da and 5368.5 Da consistent with the presence of both cytosine and guanine SNP.

Figure 2 - Representative MALDI-TOF MS spectra. Genotypes are determined by plotting peak intensity (*y*-axis) against mass (Da) (*x*-axis).

Blood Transfus 2015; 13: 78-85 DOI 10.2450/2014.0336-13



Figure 3 - Nucleotide sequencing chromatograms showing the 252C>G nucleotide polymorphism. The arrow indicates the location of cytosine (C) in IN*A/A to form proline and guanine (G) in IN*B/B to form arginine. Both C and G are co-incident for the heterozygous samples IN*A/B.

Table III - Mean fluorescence intensity (MFI) ratio of controlIn(a+b-), control In(a-b+), and two identifiedIn(a+b+) red cells measured with anti-In^a and anti-In^b. Due to the limited supply of antisera, the MFIratio values supplied in this table were taken froma single experiment only.

Sample	MFI ratio using anti-In ^a	MFI ratio using anti-In ^b	
Control In(a+b-)	4.9	1.0	
Control In(a-b+)	1.0	4.7	
Donor 1 In(a+b+)	2.7	1.8	
Donor 2 In(a+b+)	3.2	4.2	

Serology and In^a frequency in the selected population

All 151 samples were agglutinated by anti-In^b antiserum, indicating the presence of the In^b antigen.

Red blood cells of the two samples genotyped as IN^*A/B were further tested using anti-In^a antiserum. Both cells were positive for agglutination, confirming the phenotype In(a+b+). The frequency of In^a was 1.3% in this selected group.

Flow cytometry

Flow cytometry was performed with anti-In^a and anti-In^b antisera (Table III). To calculate mean fluorescence intensity (MFI) ratios, we divided the MFI value of

the sample by the MFI value of the negative control. In(a+b-) and In(a-b+) cells were used as negative controls for anti-In^b and anti-In^a flow cytometry assays, respectively. Negative controls were given an MFI ratio of 1.0. The two donor samples identified as IN*A/B gave MFI values between those of the control cells, further confirming that these donor cells express In^a and In^b antigens. The MFI ratios of the two heterozygous samples differed in their reactivity to each particular antisera, possibly because of variable expression of antigens on the cell surface. Due to the limited supply of antisera, it was not possible to repeat this assay.

Discussion

In this study on the detection of In^a blood group antigen in the Australian population, this antigen was detected in 1.3% of a cohort of donors. This rate is intermediate between the 0.1% occurrence in Caucasians⁴ and the high rate observed for Middle East groups¹⁶. Two genotyping methods, one novel PCRbased HRM assay and a commercial MassARRAY system, detected two (2/151) blood donors exhibiting the IN*A SNP along with the IN*B SNP. Serology testing using the rare anti-In^a confirmed heterozygosity for the In^a antigen in these samples. The concordance rate between genotyping systems and serotyping for the high frequency In^b was 100%.

The design of both the genotyping assays here was based on the SNP from guanine (encoding In^b) to cytosine (encoding In^a) at nucleotide 252 in exon 2 of CD44. This is the first application of HRM genotyping for Indian antigens although standard HRM analysis by amplicon melting has been applied to other blood groups²⁸. However, our experience with the HRM technique showed that standard amplicon melting alone is insufficient to detect a nucleotide change between "C" and "G". The amplicon T_{m} difference in homozygotes with class 3 (C::G) polymorphisms is very small, usually <0.4 °C as class 3 SNP are a simple inversion of a complementary base pair¹⁹. To overcome this complexity, we introduced an unlabelled probe into the reaction mix. This addition improved the specificity of our HRM genotyping assay by generating distinct melt profiles for samples homozygous for IN^*A or IN*B. The novel, closed-tube, real-time PCR method does not require post-amplification sample processing compared to IN^*A/B genotyping methods by DNA sequence typing⁹ and gel electrophoresis²⁹. The PCR UP-HRM described here is a medium-throughput assay capable of genotyping 32 samples in duplicate in 96 minutes. This assay is suitable for referral testing in specialised laboratories.

In contrast, the MassARRAY system has the advantage of being a high-throughput system capable of performing large-scale screening of up to 4,000 samples per day. The MassARRAY system which utilises MALDI-TOF MS technology was adapted through the design of primers for the IN*A and IN*B SNP. Sequenom Inc. in collaboration with the Blood Transfusion Service Zurich of the Swiss Red Cross are currently developing and evaluating a MassARRAY module for genotyping rare red blood cell polymorphisms, including IN*A and IN*B³⁰. Gassner et al.³⁰ reported a zero allele frequency for IN*A (IN*01) among 3,040 donors in this Swiss study. This is in contrast to the 1.3% found here in our population. The present study is limited by the sample size and the fact that it is confined to one Metropolitan area. Interestingly, 2% of the Australian population identify as having Indian ancestry³¹. Joshi and Vasantha described serological screening of 2,000 blood units to find two compatible In(a+b-) cells for transfusion to a patient with anti-In^b antibody³², demonstrating that serological typing to find donors with rare phenotypes is challenging when antisera are rare and in limited supply³³. As the antisera supply was limited in this study, serological typing for In^a was restricted to the two samples genotyped as IN^*A/B (by both PCR UP-HRM and MassARRAY). Serology by agglutination and

flow cytometry studies confirmed that the In^a antigen was present and the flow cytometry findings were consistent with single-dose gene expression. In future studies, flow cytometry may be preferred as less antisera is required compared to that necessary for red cell agglutination. Additionally, future studies could apply either genotyping methods described here to a larger number of samples to reliably establish the IN^*A/B allele frequency within the Australian population.

In conclusion, the frequency of ethnically-associated blood group antigens is expected to vary as the population demographic changes. The Australian Red Cross Blood Service has an active policy to recruit donors from different backgrounds to prepare to meet patients' needs in the future. Serological typing to find donors with uncommon phenotypes is challenging especially when antisera are rare and in limited supply³³. Genotyping methods will play a growing role in defining rare alleles such as IN^*A in the Australian blood donor panel and direct serology to confirm blood donors with the rare In(a+b-) phenotype. As evidenced by the finding of IN^*A/B in this sample group, PCR UP-HRM and MassARRAY are viable methods for the detection of the IN*A and IN*B alleles. Molecular typing is a useful adjunct for Indian typing when antisera are limited in supply.

Acknowledgements

The Australian Government fully funds the Australian Red Cross Blood Service for the provision of blood products and services to the Australian community. We thank Dr Melinda Dean and Dr Stacy Scott for their contribution to the flow cytometry data.

We are very grateful to the Serum, Cells, And Rare Fluids exchange programme for the gift of anti-In^b antiserum and to Joyce Poole and Nicole Thorton of the International Blood Group Reference Laboratory, Bristol, UK, for the gift of anti-In^a antiserum. We thank Andrew Davis from the Red Cell Reference Laboratory, Australian Red Cross Blood Service, Sydney, NSW, for kindly supplying the In^a cell.

This study was approved by the Australian Red Cross Blood Service Ethics Committee (reference number 2010#07).

Authorship contribution

All Authors contributed in writing and reviewing the manuscript. Genghis H. Lopez and Rhiannon S. McBean contributed equally in designing the experiments, performing molecular testing and data analysis. Darryl L. Irwin provided expert advice on the MALDI-TOF MS technology and assay design. Brett Wilson and Yew-Wah Liew performed phenotyping tests and analysed serological data. Catherine A. Hyland and Robert L. Flower supervised the study design and reviewed molecular and serological data.

Disclosure and competing interests statement

The Authors from the Australian Red Cross Blood Service have no competing interests. Darryl L. Irwin is the Applications and Technology Director of Sequenom Inc. Asia Pacific.

References

- Screaton GR, Bell MV, Jackson DG, et al. Genomic structure of DNA encoding the lymphocyte homing receptor CD44 reveals at least 12 alternatively spliced exons. Proc Natl Acad Sci USA 1992; 89: 12160-4.
- Daniels G. *Human Blood Groups*. 3rd ed. Oxford, UK: Wiley-Blackwell Publishing 2013.
- Tolg C, Hofmann M, Herrlich P, Ponta H. Splicing choice from ten variant exons establishes CD44 variability. Nucleic Acids Res 1993; 21: 1225-9.
- Reid ME, Lomas-Francis C, Olsson ML. *The Blood Group* Antigen Facts Book. 3rd ed. Elsevier Academic Press; 2012.
- 5) Spring FA, Dalchau R, Daniels GL, et al. The Ina and Inb blood group antigens are located on a glycoprotein of 80,000 MW (the CDw44 glycoprotein) whose expression is influenced by the In(Lu) gene. Immunology 1988; 64: 37-43.
- Badakere SS, Joshi SR, Bhatia HM, et al. Evidence for a new blood group antigen in the Indian population (a preliminary report). Indian J Med Res 1973; 61: 563.
- Badakere SS, Parab BB, Bhatia HM. Further observations on the Ina (Indian) antigen in Indian populations. Vox Sang 1974; 26: 400-3.
- Giles CM. Antithetical relationship of anti-In-a with the Salis antibody. Vox Sang 1975; 29: 73-6.
- Poole J, Tilley L, Warke N, et al. Two missense mutations in the CD44 gene encode two new antigens of the Indian blood group system. Transfusion 2007; 47: 1306-11.
- Telen MJ, Udani M, Washington MK, et al. A blood grouprelated polymorphism of CD44 abolishes a hyaluronan-binding consensus sequence without preventing hyaluronan Bbnding. J Biol Chem 1996; 271: 7147-53.
- Longster GH, North DI, Robinson EA. Four further examples of anti-Inb detected during pregnancy. Clin Lab Haematol 1981; 3: 351-6.
- Bhatia H. Studies on the blood group antigen Ina. Immunol Commun 1980; 9: 203.
- 13) Ferguson D, Gaal H. Some observations on the Inb antigen and evidence that anti-Inb causes accelerated destruction of radiolabeled red cells. Transfusion 1988; 28: 479-82.
- Anstee DJ. Red cell genotyping and the future of pretransfusion testing. Blood 2009; 114: 248-56.
- Joshi SR. Immediate haemolytic transfusion reaction due to anti-Inb. Vox Sang 1992; 63: 232-3.
- Badakere SS, Vasantha K, Bhatia HM, et al. High frequency of Ina antigen among Iranians and Arabs. Hum Hered 1980; 30: 262-3.
- Bhatia HM, Badakere SS, Mokashi SA, Parab BB. Studies on the blood group antigen Ina. Immunol Commun 1980; 9: 203-15.
- Montgomery J, Wittwer CT, Palais R, Zhou L. Simultaneous mutation scanning and genotyping by high-resolution DNA melting analysis. Nat Protoc 2007; 2: 59-66.
- Liew M, Pryor R, Palais R, et al. Genotyping of singlenucleotide polymorphisms by high-resolution melting of small amplicons. Clin Chem 2004; 50: 1156-64.

- 20) Reed GH, Kent JO, Wittwer CT. High-resolution DNA melting analysis for simple and efficient molecular diagnostics. Pharmacogenomics 2007; 8: 597-608.
- 21) Zhou L, Myers AN, Vandersteen JG, et al. Closed-tube genotyping with unlabeled oligonucleotide probes and a saturating DNA dye. Clin Chem 2004; 50: 1328-35.
- 22) Erali M, Voelkerding KV, Wittwer CT. High resolution melting applications for clinical laboratory medicine. Exp Mol Pathol 2008; 85: 50-8.
- 23) Liew M, Seipp M, Durtschi J, et al. Closed-tube SNP genotyping without labeled probes/a comparison between unlabeled probe and amplicon melting. Am J Clin Pathol 2007; **127**: 341-8.
- 24) Dames S, Margraf RL, Pattison DC, et al. Characterization of aberrant melting peaks in unlabeled probe assays. J Mol Diagn 2007; 9: 290-6.
- 25) Sauer S, Gut IG. Genotyping single-nucleotide polymorphisms by matrix-assisted laser-desorption/ionization time-of-flight mass spectrometry. J Chromatogr B 2002; 782: 73-87.
- 26) Tost J, Gut IG. Genotyping single nucleotide polymorphisms by MALDI mass spectrometry in clinical applications. Clin Biochem 2005; 38: 335-50.
- 27) Garritsen HS, Fan AX, Bosse N, et al. Matrix-assisted laser desorption/ionization time-of-flight mass spectrometry for genotyping of human platelet-specific antigens. Transfusion 2009; 49: 252-8.
- 28) Tanaka M, Takahahi J, Hirayama F, Tani Y. High-resolution melting analysis for genotyping Duffy, Kidd and Diego blood group antigens. Leg Med (Tokyo) 2011; 13: 1-6.
- 29) Jungbauer C, Hobel CM, Schwartz DW, Mayr WR. Highthroughput multiplex PCR genotyping for 35 red blood cell antigens in blood donors. Vox Sang 2012; 102: 234-42.
- 30) Gassner C, Meyer S, Frey BM, Vollmert C. Matrix-assisted laser desorption/ionisation, time-of-flight mass spectrometry-based blood group genotyping--the alternative approach. Transfus Med Rev 2013; 27: 2-9.
- 31) Cultural Diversity in Australia, Reflecting a Nation: Stories from the 2011 Census, Australian Bureau of Statistics. Available at: http://www.abs.gov.au. Accessed on 16/12/2013.
- 32) Joshi SR, Vasantha K. A profile of rare bloods in India and its impact in blood transfusion service. Asian J Transfus Sci 2012; 6: 42-3.
- 33) Reid ME. DNA analysis to find rare blood donors when antisera is not available. Vox Sang 2002; 83 (Suppl 1): 91-3.

Arrived: 19 December 2013 - Revision accepted: 27 February 2014 Correspondence: Genghis H. Lopez Research and Development Laboratory Australian Red Cross Blood Service 44, Musk Avenue Kelvin Grove, Queensland 4059, Australia e-mail: GLopez@redcrossblood.org.au

Blood Transfus 2015; 13: 78-85 DOI 10.2450/2014.0336-13